



# Mouse IL-1 $\beta$ ELISA Kit

**Vertrieb:**

**L O X O GmbH** Immunbiologie Biochemie, Produkte und Systeme  
Postfach 11 30 69215 Dossenheim  
**Telefon** +49 (0) 62 21 - 86 80 23 **FAX** +49 (0) 62 21 - 86 80 255  
**E-Mail:** info@loxo.de **Internet:** www.loxo.de

Assaypro LLC  
30 Triad South Drive  
St. Charles, MO 63304  
T (636) 447-9175  
F (636) 447-9475  
[www.assaypro.com](http://www.assaypro.com)

**Hinweis/Note:**

Der Packungsbeileger dient nur als erste Information.  
Der relevante Packungsbeileger liegt der Ware bei.

The datasheet is only a first information.  
The relevant datasheet is included with the product.

For any questions regarding troubleshooting or performing the assay, please contact our support team at [support@assaypro.com](mailto:support@assaypro.com).

Thank you for choosing Assaypro.

## Assay Summary

Add 50  $\mu$ l of standard/samples per well.  
Incubate 2 hours.



Wash, then add 50  $\mu$ l of  
biotinylated antibody per well.  
Incubate 2 hours.



Wash, then add 50  $\mu$ l of SP per well.  
Incubate 30 minutes.



Wash, then add 50  $\mu$ l of  
Chromogen Substrate per well.  
Incubate 25 minutes.



Add 50  $\mu$ l of Stop Solution per well.  
Read at 450 nm immediately.

	1	2	3	4	5	6	7	8	9	10	11	12
A												
B												
C												
D												
E												
F												
G												
H												

**Assay Template**



# **AssayMax Mouse Interleukin-1 $\beta$ (IL-1 $\beta$ ) ELISA Kit**

Catalog No. EMI2200-1  
Sample Insert/Reference Only

## **Introduction**

Interleukin-1 $\beta$  (IL-1 $\beta$ ) has a wide spectrum of inflammatory, metabolic, haemopoietic, and immunological properties (1). IL-1 $\beta$  plays a significant role in hippocampal synaptic function (2) and is a potential genetic marker as indicator of gastric cancer risk (3). High plasma level of interleukin-1 $\beta$  is associated with rheumatoid and osteoarthritic joint disease (4), infectious gastroenteritis (5), neurodegeneration (6), and breast cancer (7). High gingival crevicular fluid levels of IL-1 $\beta$  are related to type 2 diabetes (8).

## **Principle of the Assay**

The AssayMax Mouse Interleukin-1 $\beta$  ELISA kit is designed for detection of mouse IL-1 $\beta$  in plasma, serum, and cell culture supernatants. This assay employs a quantitative sandwich enzyme immunoassay technique, which measures IL-1 $\beta$  in less than 5 hours. A murine monoclonal antibody specific for IL-1 $\beta$  has been pre-coated onto a microplate. IL-1 $\beta$  in standards and samples is sandwiched by the immobilized antibody and a biotinylated polyclonal antibody specific for IL-1 $\beta$ , which is recognized by a streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

## **Caution and Warning**

- **Prepare all reagents (working diluent buffer, wash buffer, standards, biotinylated antibody, and SP conjugate) as instructed, prior to running the assay.**
- **Prepare all samples prior to running the assay. The dilution factors for the samples are suggested in this protocol. However, the user should determine the optimal dilution factor.**
- **Spin down the SP conjugate vial and the biotinylated antibody vial before opening and using contents.**
- This kit is for research use only.
- The kit should not be used beyond the expiration date.

- The Stop Solution is an acidic solution.

## Reagents

- **Mouse IL-1 $\beta$  Microplate:** A 96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody against mouse IL-1 $\beta$ .
- **Sealing Tapes:** Each kit contains 3 precut, pressure sensitive sealing tapes that can be cut to fit the format of the individual assay.
- **Mouse IL-1 $\beta$  Standard:** Mouse IL-1 $\beta$  in a buffered protein base (2 ng, lyophilized).
- **Biotinylated Mouse IL-1 $\beta$  Antibody (100x):** A 100-fold biotinylated polyclonal antibody against mouse IL-1 $\beta$  (80  $\mu$ l).
- **MIX Diluent Concentrate (10x):** A 10-fold concentrated buffered protein base (30 ml).
- **Wash Buffer Concentrate (20x):** A 20-fold concentrated buffered surfactant (30 ml, 2 bottles).
- **Streptavidin-Peroxidase Conjugate (SP Conjugate):** A 100-fold concentrate (80  $\mu$ l).
- **Chromogen Substrate:** A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml).
- **Stop Solution:** A 0.5 N hydrochloric acid to stop the chromogen substrate reaction (12 ml).

## Storage Condition

- Store components of the kit at 2-8°C or -20°C upon arrival up to the expiration date.
- Store SP Conjugate and biotinylated antibody at -20°C.
- Store Microplate, Diluent Concentrate (10x), Wash Buffer, Stop Solution, and Chromogen Substrate at 2-8°C.
- Unused microplate wells may be returned to the foil pouch with the desiccant packs and resealed. May be stored for up to 1 month in a vacuum desiccator.
- Diluent (1x) may be stored for up to 1 month at 2-8°C.
- Store standard at 2-8°C before reconstituting with diluent and at -20°C after reconstituting with diluent.

## Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm.
- Pipettes (1-20  $\mu$ l, 20-200  $\mu$ l, 200-1000  $\mu$ l and multiple channel).
- Deionized or distilled reagent grade water.

## Sample Collection, Preparation and Storage

- **Plasma:** Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 3000 x *g* for 10 minutes and assay. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles. (EDTA or Heparin can also be used as an anticoagulant.)
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 3000 x *g* for 10 minutes. Remove serum and assay. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- **Cell Culture Supernatants:** Centrifuge cell culture media at 3000 x *g* for 10 minutes to remove debris. Collect supernatants and assay. Store the remaining samples at -20°C or below. Avoid repeated freeze-thaw cycles.

## Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use.
- **MIX Diluent Concentrate (10x):** If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the MIX Diluent Concentrate 1:10 with reagent grade water. Store for up to 1 month at 2-8°C.
- **Standard Curve:** Reconstitute the 2 ng of Mouse IL-1 $\beta$  Standard with 2 ml of MIX Diluent to generate a standard solution of 1 ng/ml. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially diluting the IL-1 $\beta$  standard solution (1 ng/ml) 1:2 with equal volume of MIX Diluent to produce 0.5, 0.25, 0.125, 0.063, 0.031, and 0.016 ng/ml solutions. MIX Diluent serves as the zero standard (0 ng/ml). Any remaining solution should be frozen at -20°C.

Standard Point	Dilution	[Mouse IL-1 $\beta$ ] (ng/ml)
P1	1 part Standard (1 ng/ml)	1.000
P2	1 part P1 + 1 part MIX Diluent	0.500
P3	1 part P2 + 1 part MIX Diluent	0.250
P4	1 part P3 + 1 part MIX Diluent	0.125
P5	1 part P4 + 1 part MIX Diluent	0.063
P6	1 part P5 + 1 part MIX Diluent	0.031
P7	1 part P6 + 1 part MIX Diluent	0.016
P8	MIX Diluent	0.000

- **Biotinylated Mouse IL-1 $\beta$  Antibody (100x):** Spin down the antibody briefly and dilute the desired amount of the antibody 1:100 with MIX Diluent. Any remaining solution should be frozen at -20°C.
- **Wash Buffer Concentrate (20x):** If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute Wash Buffer Concentrate 1:20 with reagent grade water.
- **SP Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with MIX Diluent. Any remaining solution should be frozen at -20°C.

## Assay Procedure

- Prepare all reagents, working standards and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-30°C).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccants inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50  $\mu$ l of Mouse IL-1 $\beta$  Standard or sample per well. Cover wells with a sealing tape and incubate for 2 hours. Start the timer after the last sample addition.
- Wash five times with 200  $\mu$ l of Wash Buffer manually. Invert the plate each time and decant the contents; hit 4-5 times on absorbent material to completely remove the liquid. If using a machine, wash six times with 300  $\mu$ l of Wash Buffer and then invert the plate, decanting the contents; hit 4-5 times on absorbent material to completely remove the liquid.
- Add 50  $\mu$ l of Biotinylated Mouse IL-1 $\beta$  Antibody to each well and incubate for 2 hours.
- Wash the microplate as described above.
- Add 50  $\mu$ l of Streptavidin-Peroxidase Conjugate per well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash the microplate as described above.
- Add 50  $\mu$ l of Chromogen Substrate per well and incubate for about 25 minutes or until the optimal blue color density develops. Gently tap the plate to ensure thorough mixing and break the bubbles in the well with pipette tip.
- Add 50  $\mu$ l of Stop Solution to each well. The color will change from blue to yellow.
- Read the absorbance on a microplate reader at a wavelength of 450 nm **immediately**. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some




unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings.

## Data Analysis

- Calculate the mean value of the triplicate readings for each standard and sample.
- To generate a standard curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using log-log or four-parameter logistic curve-fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

## Standard Curve

- The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.



## Performance Characteristics

- The minimum detectable level of IL-1beta is typically ~0.016 ng/ml.
- Intra-assay and inter-assay coefficients of variation were 4.3 % and 7.5 % respectively.

## Cross-Reactivity

- No significant cross-reactivity or interference was observed.

## Recovery

<b>Standard Added Value</b>	0.05 – 0.5 ng/ml
<b>Recovery %</b>	88-112 %
<b>Average Recovery %</b>	98 %

## References

- (1) Palmi M *et al.* (2002) *Mol Neurobiol.* 25(2): 133-47
- (2) Lynch MA (2002) *Vitam Horm.* 64:185-219
- (3) Troost E *et al.* (2003) *Can J Gastroenterol.* 17 Suppl B: 8B-12B
- (4) Fan Z *et al.* (2004) *Cytokine* 28(1): 17-24
- (5) Enocksson A *et al.* (2004) *Clin Diagn Lab Immunol.* 11(2): 250-4
- (6) Allan SM *et al.* (2005) *Nat Rev Immunol* 5(8): 629-40
- (7) Mettler L *et al.* (2004) *Clin Exp Obstet Gynecol.* 31(1): 20-2
- (8) Engebretson SP *et al.* (2004) *J Periodontol.* 75(9): 1203-8

Version 1.5R1

## Related products

- EI2200-1 AssayMax Human Interleukin-1 $\beta$  ELISA Kit (Plasma, Serum and Cell Culture Supernatants samples)