



Mouse IGF-1 ELISA Kit

Vertrieb:

L O X O GmbH Immunbiologie Biochemie, Produkte und Systeme
Postfach 11 30 69215 Dossenheim
Telefon +49 (0) 62 21 - 86 80 23 **FAX** +49 (0) 62 21 - 86 80 255
E-Mail: info@loxo.de **Internet:** www.loxo.de

Assaypro LLC
30 Triad South Drive
St. Charles, MO 63304
T (636) 447-9175
F (636) 447-9475
www.assaypro.com

Hinweis/Note:

Der Packungsbeileger dient nur als erste Information.
Der relevante Packungsbeileger liegt der Ware bei.

The datasheet is only a first information.
The relevant datasheet is included with the product.


For any questions regarding troubleshooting or performing the assay, please contact our support team at support@assaypro.com.

Thank you for choosing Assaypro.

Assay Summary

Add 50 μ l of standard/samples per well.

Incubate 2 hours.




Wash, then add 50 μ l of
biotinylated antibody per well.

Incubate 2 hours.




Wash, then add 50 μ l of SP per well.

Incubate 30 minutes.



Wash, then add 50 μ l of
Chromogen Substrate per well.

Incubate 12 minutes.



Add 50 μ l of Stop Solution per well.

Read at 450 nm immediately.

AssayMax Mouse Insulin-like Growth Factor 1 (IGF-1) ELISA Kit

Catalog No. EMI1001-1
Sample Insert/Reference Only

Introduction

Insulin-like growth factor 1 (IGF-1) is a 70 amino acid polypeptide protein hormone with molecular mass of 7.65 kDa (1). IGF-1 is produced primarily by the liver in response to the stimulation of growth hormone. It is transported in plasma bound to different forms of IGF-1 binding proteins (2). It also binds to specific IGF-1 tyrosine kinase receptor and the insulin receptor. Inhibition IGF-1 receptor reduces pancreatic cancer growth and angiogenesis (3). IGF-I regulates cellular proliferation, differentiation, apoptosis, and amyloid precursor protein family (4 - 5). It may be important in the pathophysiological processes underlying chronic disease, including type 2 diabetes mellitus, coronary heart disease, cancer, and Alzheimer's disease (6 - 8). Increased levels of IGF lead to an increased risk of cancer (9). IGF-I stimulates osteoblast proliferation, bone formation, and increases bone volume (10). It is a potent neurotrophic as well as a neuroprotective factor found in the central and the peripheral nervous systems of the brain (11).

Principle of the Assay

The AssayMax Mouse IGF-1 ELISA kit is designed for detection of mouse IGF-1 in plasma, serum, and cell culture supernatants. This assay employs a quantitative sandwich enzyme immunoassay technique, which measures mouse IGF-1 in less than 5 hours. A polyclonal antibody specific for mouse IGF-1 has been pre-coated onto a microplate. Mouse IGF-1 in standards and samples is sandwiched by the immobilized antibody and biotinylated polyclonal antibody specific for mouse IGF-1, which is recognized by a streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

Caution and Warning

- **Prepare all reagents (working diluent buffer, wash buffer, standards, biotinylated antibody, and SP conjugate) as instructed, prior to running the assay.**

- **Prepare all samples prior to running the assay. The dilution factors for the samples are suggested in this protocol. However, the user should determine the optimal dilution factor.**
- **Spin down the SP conjugate vial and the biotinylated antibody vial before opening and using contents.**
- This kit is for research use only.
- The kit should not be used beyond the expiration date.
- The Stop Solution is an acidic solution.

Reagents

- **Mouse IGF-1 Microplate:** A 96 well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against mouse IGF-1.
- **Sealing Tapes:** Each kit contains 3 precut, pressure sensitive sealing tapes, which can be cut to fit the format of the individual assay.
- **Mouse IGF-1 Standard:** Mouse IGF-1 in a buffered protein base (40 ng, lyophilized).
- **Biotinylated Mouse IGF-1 Antibody (50x):** A 50-fold concentrated biotinylated polyclonal antibody against IGF-1 (140 μ l).
- **Streptavidin-Peroxidase Conjugate (SP Conjugate):** A 100-fold concentrate (80 μ l).
- **MIX Diluent Concentrate (10x):** A 10-fold concentrated buffered protein base (30 ml).
- **Wash Buffer Concentrate (20x):** A 20-fold concentrated buffered surfactant (30 ml, 2 bottles).
- **Chromogen Substrate:** A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml).
- **Stop Solution:** A 0.5 N hydrochloric acid to stop the chromogen substrate reaction (12 ml).

Storage Condition

- Store components of the kit at 2-8°C or -20°C upon arrival up to the expiration date.
- Store SP Conjugate and biotinylated antibody at -20°C.
- Store Microplate, Diluent Concentrate (10x), Wash Buffer, Stop Solution, and Chromogen Substrate at 2-8°C.
- Unused microplate wells may be returned to the foil pouch with the desiccant packs and resealed. May be stored for up to 1 month in a vacuum desiccator.
- Diluent (1x) may be stored for up to 1 month at 2-8°C.
- Store standard at 2-8°C before reconstituting with diluent and at -20°C after reconstituting with diluent.

Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm.
- Pipettes (1-20 μ l, 20-200 μ l, 200-1000 μ l and multiple channel).
- Deionized or distilled reagent grade water.

Sample Collection and Storage

- **Plasma:** Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 3000 x *g* for 10 minutes. Dilute samples 1:250 into MIX Diluent and assay. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 3000 x *g* for 10 minutes and remove serum. Dilute samples 1:250 into MIX Diluent and assay. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- **Cell Culture Supernatants:** Centrifuge cell culture media at 3000 x *g* for 10 minutes to remove debris. Collect supernatants and assay. Store samples at -20°C or below. Avoid repeated freeze-thaw cycles.

Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use.
- **MIX Diluent Concentrate (10x):** If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the MIX Diluent Concentrate 1:10 with reagent grade water. Store for up to 1 month at 2-8°C.
- **Mouse IGF-1 Standard:** Reconstitute the 40 ng of Mouse IGF-1 Standard with 4 ml of MIX Diluent to generate a stock solution of 10 ng/ml. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially diluting the standard solution (10 ng/ml) 1:2 with equal volume of MIX Diluent to produce 5, 2.5, 1.25, 0.625, 0.313 and 0.156 ng/ml solutions. MIX Diluent serves as the zero standard (0 ng/ml). Any remaining solution should be frozen at -20°C and used within the next 30 days.

Standard Point	Dilution	[Mouse IGF-1] (ng/ml)
P1	Standard (10 ng/ml)	10.00
P2	1 part P1 + 1 part MIX Diluent	5.000
P3	1 part P2 + 1 part MIX Diluent	2.500
P4	1 part P3 + 1 part MIX Diluent	1.250
P5	1 part P4 + 1 part MIX Diluent	0.625
P6	1 part P5 + 1 part MIX Diluent	0.313
P7	1 part P6 + 1 part MIX Diluent	0.156
P8	MIX Diluent	0.000

- **Biotinylated Mouse IGF-1 Antibody (50x):** Spin down the antibody briefly and dilute the desired amount of the antibody 1:50 with MIX Diluent. Any remaining solution should be frozen at -20°C.
- **Wash Buffer Concentrate (20x):** If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the Wash Buffer Concentrate 1:20 with reagent grade water.
- **Streptavidin-Peroxidase Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with MIX Diluent. Any remaining solution should be frozen at -20°C.

Assay Procedure

- Prepare all reagents, working standards and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-30°C).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccants inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50 µl of Mouse IGF-1 Standard or sample per well, and cover wells and incubate for 2 hours. Start the timer after the last sample addition.
- Wash five times with 200 µl of Wash Buffer manually. Invert the plate each time and decant the contents; hit 4-5 times on absorbent material to completely remove the liquid. If using a machine wash, six times with 300 µl of Wash Buffer and then invert the plate, decanting the contents; hit 4-5 times on absorbent material to completely remove the liquid.
- Add 50 µl of Biotinylated Mouse IGF-1 Antibody to each well and incubate for 2 hours.
- Wash the microplate as described above.
- Add 50 µl of Streptavidin-Peroxidase Conjugate per well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash the microplate as described above.


- Add 50 μ l of Chromogen Substrate per well and incubate for about 12 minutes or till the optimal blue color density develops. Gently tap plate to ensure thorough mixing and break the bubbles in the well with pipette tip.
- Add 50 μ l of Stop Solution to each well. The color will change from blue to yellow.
- Read the absorbance on a microplate reader at a wavelength of 450 nm **immediately**. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings.

Data Analysis

- Calculate the mean value of the duplicate or triplicate readings for each standard and sample.
- To generate a standard curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using 4-parameter or log-log logistic curve-fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

Standard Curve

- The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.



Performance Characteristics

- The minimum detectable dose of Mouse IGF-1 is typically ~ 0.1 ng/ml.
- Intra-assay and inter-assay coefficients of variation were 4.9 % and 7.1% respectively.

Linearity

Sample Dilution	Average Percentage of Expected Value	
	Plasma	Serum
1:125	94%	97%
1:250	100%	99%
1:500	104%	104%

Recovery

Standard Added Value	0.3 – 3 ng/ml
Recovery %	86-112 %
Average Recovery %	98.5%

Cross-Reactivity

Species	% Cross Reactivity
Canine	None
Monkey	None
Mouse	100%
Rat	<70%
Swine	<5%
Bovine	None
Rabbit	None
Human	<2%

References

- (1) Sandberg-Nordqvist AC *et al.* (1993) *Cancer Res.* 53: 2475-2478
- (2) Jones JJ and Clemmons DR (1995) *Endocr. Rev.* 16: 3-34
- (3) Moser C. *et al.* (2008) *Eur. J. Cancer* 44: 1577-1586
- (4) Remacle-Bonnet MM *et al.* (2000) *Cancer Res.* 60: 2007-2017
- (5) Adlerz L *et al.* (2007) *J. Biol. Chem.* 282: 10203 -10209
- (6) Sandhu MS (2005) *Endocr. Dev.* 9:44-54
- (7) Sima AA and Li ZG (2006) *Rev. Diabet. Stud.* 3: 161-168
- (8) Butovsky O *et al.* (2006) *Proc. Natl. Acad. Sci. U.S.A.* 103: 11784-11789
- (9) Jo H *et al.* (2008) *Mol. Cell Biol.* 28: 4285-4299
- (10) Spencer EM *et al.* (1991) *Bone* 12: 21-26
- (11) Gomez JM (2008) *Endocr. Metab. Immune Disord Drug Targets* 8: 143-151

Version 2.2

Related products

- EI1001-1 AssayMax Human IGF-1 ELISA Kit (Plasma, Serum, and Cell Culture samples)