



Human HSP60 ELISA Kit

Vertrieb:

L O X O GmbH Immunbiologie Biochemie, Produkte und Systeme
Postfach 11 30 69215 Dossenheim
Telefon +49 (0) 62 21 - 86 80 23 **FAX** +49 (0) 62 21 - 86 80 255
E-Mail: info@loxo.de **Internet:** www.loxo.de

Assaypro LLC
30 Triad South Drive
St. Charles, MO 63304
T (636) 447-9175
F (636) 447-9475
www.assaypro.com

Hinweis/Note:

Der Packungsbeileger dient nur als erste Information.
Der relevante Packungsbeileger liegt der Ware bei.

The datasheet is only a first information.
The relevant datasheet is included with the product.

For any questions regarding troubleshooting or performing the assay, please contact our support team at support@assaypro.com.

Thank you for choosing Assaypro.

Assay Summary

Add 50 μ l of Standard/ Sample per well.
Incubate 2 hours.




Wash, then add
50 μ l of Biotinylated Antibody per well.
Incubate 2 hours.



Wash, then add
50 μ l of SP Conjugate per well.
Incubate 30 minutes.



Wash, then add
50 μ l of Chromogen Substrate per well.
Incubate 30 minutes.



Add 50 μ l of Stop Solution per well.
Read at 450 nm immediately.

AssayMax Human Heat Shock Protein 60 (HSP60) ELISA Kit

Catalog No. EH5505-1
Sample Insert/Reference Only

Introduction

Heat shock protein of 60 kDa (HSP60) is a mitochondrial chaperonin involved in folding, assembly, and transport of newly imported protein from cytoplasm into mitochondria in an ATP-mediated reaction (1-3). Human HSP60 contains 573 amino acids and is related to the bacteria groEL protein. HSP60 is located in the mitochondria and cytoplasm, the cell surface, the extracellular space, and the peripheral blood (4, 5). Under dehydration conditions, the cytoplasmic HSP60 is quickly imported into the mitochondria by cytoplasmic HSP70 (6). Extracellular HSP60 mediates apoptosis via Toll-like receptors in heart failure (7). HSP60 plays a role in myelinogenesis and neurodegeneration: Its defect can cause neurodegenerative pathologies (8). It has been suggested that HSP60 is linked to diabetes, stress response, cancer, and immunological disorders. Since it participates in cell signaling and key pathways and actively secreted by human tumor cells, HSP60 could be used for cancer diagnosis and therapy (9).

Principle of the Assay

The AssayMax Human Heat Shock Protein 60 (HSP60) ELISA (Enzyme-Linked Immunosorbent Assay) kit is designed for detection of human HSP60 in plasma, serum, cell culture lysates, and tissue samples. This assay employs a quantitative sandwich enzyme immunoassay technique that measures human HSP60 in 5 hours. A polyclonal antibody specific for human HSP60 has been pre-coated onto a 96-well microplate with removable strips. HSP60 in standards and samples is sandwiched by the immobilized antibody and biotinylated polyclonal antibody specific for HSP60, which is recognized by a streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

Caution and Warning

- Prepare all reagents (working diluent buffer, wash buffer, standard, biotinylated antibody, and SP conjugate) as instructed, prior to running the assay.
- Prepare all samples prior to running the assay. The dilution factors for the samples are suggested in this protocol. However, the user should determine the optimal dilution factor.
- Spin down the SP conjugate vial and the biotinylated antibody vial before opening and using contents.
- This kit is for research use only.
- The kit should not be used beyond the expiration date.
- The Stop Solution is an acidic solution.

Reagents

- **Human HSP60 Microplate:** A 96-well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against human HSP60.
- **Sealing Tapes:** Each kit contains 3 precut, pressure sensitive sealing tapes that can be cut to fit the format of the individual assay.
- **Human HSP60 Standard:** Human HSP60 in a buffered protein base (120 ng, lyophilized, 2 bottles).
- **Biotinylated Human HSP60 Antibody (70x):** A 70-fold concentrated biotinylated polyclonal antibody against HSP60 (105 μ l).
- **EIA Diluent Concentrate (10x):** A 10-fold concentrated buffered protein base (20 ml).
- **Wash Buffer Concentrate (20x):** A 20-fold concentrated buffered surfactant (30 ml, 2 bottles).
- **Streptavidin-Peroxidase Conjugate (SP Conjugate):** A 100-fold concentrated (80 μ l).
- **Chromogen Substrate:** A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml).
- **Stop Solution:** A 0.5 N hydrochloric acid to stop the chromogen substrate reaction (12 ml).

Storage Condition

- Upon arrival, immediately store components of the kit at recommended temperatures up to the expiration date.
- Store SP Conjugate and Biotinylated Antibody at -20°C.
- Store Microplate, Diluent Concentrate (10x), Wash Buffer, Stop Solution, and Chromogen Substrate at 2-8°C.

- Unused microplate wells may be returned to the foil pouch with the desiccant packs and resealed. May be stored for up to 30 days in a vacuum desiccator.
- Diluent (1x) may be stored for up to 30 days at 2-8°C.
- Store Standard at 2-8°C before reconstituting with Diluent. Fresh standard should be reconstituted the day the assay is run.

Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm.
- Pipettes (1-20 μ l, 20-200 μ l, 200-1000 μ l, and multiple channel).
- Deionized or distilled reagent grade water.

Sample Collection, Preparation and Storage

- **Plasma:** Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 3000 x *g* for 10 minutes and assay. Store samples at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles (EDTA or Heparin can also be used as an anticoagulant).
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 3000 x *g* for 10 minutes. Collect the sample and assay. Store samples at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- **Cell Culture Lysates:** Place the cell culture dish in ice and wash the cells with ice-cold PBS. Drain the PBS, then add ice-cold lysis buffer (20 mM Tris-HCl (pH 7.5), 150 mM NaCl, 1 mM Na₂EDTA, 1 mM EGTA, 1% Triton, 0.1 mM PMSF, 1 μ g/ml leupeptin, 1 μ g/mL aprotinin, and 1 μ g/mL pepstatin). Scrape adherent cells off the dish and transfer the cell suspension into a pre-cooled microfuge tube. Maintain constant agitation for 30 minutes at 4°C. Centrifuge in a microcentrifuge at 4°C. Collect fresh cell lysates. Use undiluted samples or 1:2 diluted samples with EIA Diluent and assay. The undiluted samples can be stored at -20°C or below.
- **Tissue:** Extract tissue samples with 50 mM phosphate-buffered saline (pH7.4) containing 1% Triton X-100 and centrifuge at 14000 x *g* for 20 minutes. Collect the supernatant and measure the protein concentration. Use undiluted samples or 1:2 diluted samples with EIA Diluent and assay. The undiluted samples can be stored at -20°C or below.

Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use.

- **EIA Diluent Concentrate (10x):** If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the EIA Diluent Concentrate 1:10 with reagent grade water. Store for up to 30 days at 2-8°C.
- **Standard Curve:** Reconstitute the 120 ng of Human HSP60 Standard with 1.5 ml of EIA Diluent to generate an 80 ng/ml standard solution. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially diluting the standard solution (80 ng/ml) 1:2 with EIA Diluent to produce 40, 20, 10, 5, and 2.5 ng/ml solutions. EIA Diluent serves as the zero standard (0 ng/ml). Any remaining solution should be discarded. **Fresh standard should be reconstituted the day the assay is run.**

Standard Point	Dilution	[HSP60] (ng/ml)
P1	Standard (80 ng/ml)	80.00
P2	1 part P1 + 1 part EIA Diluent	40.00
P3	1 part P2 + 1 part EIA Diluent	20.00
P4	1 part P3 + 1 part EIA Diluent	10.00
P5	1 part P4 + 1 part EIA Diluent	5.000
P6	1 part P5 + 1 part EIA Diluent	2.500
P7	EIA Diluent	0.000

- **Biotinylated Human HSP60 Antibody (70x):** Spin down the antibody briefly and dilute the desired amount of the antibody 1:70 with EIA Diluent. Any remaining solution should be frozen at -20°C.
- **Wash Buffer Concentrate (20x):** If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the Wash Buffer Concentrate 1:20 with reagent grade water.
- **SP Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with EIA Diluent. Any remaining solution should be frozen at -20°C.

Assay Procedure

- Prepare all reagents, standard solutions, and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-25°C).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccants inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50 µl of Human HSP60 Standard or sample per well. Cover wells with a sealing tape and incubate for 2 hours. Start the timer after the last addition.

- Wash five times with 200 μ l of Wash Buffer manually. Invert the plate each time and decant the contents; hit 4-5 times on absorbent material to completely remove the liquid. If using a machine, wash six times with 300 μ l of Wash Buffer and then invert the plate, decanting the contents; hit 4-5 times on absorbent material to completely remove the liquid.
- Add 50 μ l of Biotinylated Human HSP60 Antibody to each well and incubate for 2 hours.
- Wash the microplate as described above.
- Add 50 μ l of Streptavidin-Peroxidase Conjugate to each well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash the microplate as described above.
- Add 50 μ l of Chromogen Substrate per well and incubate for about 30 minutes or till the optimal blue color density develops. Gently tap plate to ensure thorough mixing and break the bubbles in the well with pipette tip.
- Add 50 μ l of Stop Solution to each well. The color will change from blue to yellow.
- Read the absorbance on a microplate reader at a wavelength of 450 nm **immediately**. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings.


Data Analysis

- Calculate the mean value of the duplicate or triplicate readings for each standard and sample.
- To generate a standard curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using log-log or four-parameter logistic curve-fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

Standard Curve

- The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.

Human HSP60 Standard Curve



Performance Characteristics

- The minimum detectable dose of HSP60 is typically ~ 2 ng/ml.
- Intra-assay and inter-assay coefficients of variation were 4.9% and 7.0% respectively.

Recovery

Standard Added Value	5 – 40 ng/ml
Recovery %	85 – 109%
Average Recovery %	97%

Cross-Reactivity

Species	% Cross Reactivity
Bovine	None
Human	100%
Mouse	None
Rat	None
Swine	None
Canine	None
Rabbit	None
Monkey	None

References

- (1) Cheng MY *et al.* (1990) *Nature* 348:455-458
- (2) Goloubinoff P *et al.* (1989) *Nature* 342:884-889
- (3) Ostermann J *et al.* (1989) *Nature* 341:125-130
- (4) Reading DS *et al.* (1989) *Nature* 337:655-659
- (5) Cappello F *et al.* (2008) *Cancer Biol. Ther.* 7:801-809
- (6) Itoh H *et al.* (2002) *Eur J. Biochem.* 269:5931-5938
- (7) Kim SC *et al.* (2009) *Circ. Res.* 105:1186-1195
- (8) Magen D *et al.* (2008) *Am J Hum Genet.* 83(1):30-42
- (9) Merendino AM *et al.* (2010) *PLoS One* 5(2):e9247

Version 1.3

Related Products

- EH5001-1 AssayMax Human Heat Shock Protein 27 (HSP27) ELISA Kit (Plasma, Serum, Milk, Cell Culture Lysates, and Tissue Extract samples)
- EH5202-1 AssayMax Human Heat Shock Protein 47 (HSP47) ELISA Kit (Plasma, Serum, Milk, Cell Culture Lysates, and Tissue Extract samples)